1997—Subsec. (f)(1)(B)(iii). Pub. L. 105–115, § 410(a), added cl. (iii).

Subsec. (g)(6), (7). Pub. L. 105–115, § 201(a), added pars. (6) and (7).

Subsec. (h)(4). Pub. L. 105–115, § 216(a)(1), amended par. (4) generally. Prior to amendment, par. (4) related to premarket approval of devices.

Subsec. (*l*). Pub. L. 105–115, § 125(b)(2)(E), struck out ''or antibiotic drugs'' after ''new drugs'' in heading.

Subsec. (*l*)(4). Pub. L. 105–115, § 125(b)(2)(E), struck out par. (4) which read as follows: ''Any device intended for human use which on the enactment date was subject to the requirements of section 357 of this title shall be subject to such requirements as follows:

''(A) In the case of such a device which is classified into class I, such requirements shall apply to such device until the effective date of the regulation classifying the device into such class.

''(B) In the case of such a device which is classified into class II, such requirements shall apply to such device until the effective date of a performance standard applicable to the device under section 360d of this title.

''(C) In the case of such a device which is classified into class III, such requirements shall apply to such device until the date on which the device is required to have in effect an approved application under section 360e of this title.''

Subsec. (m)(2). Pub. L. 105–115, § 203(1), inserted at end ''The request shall be in the form of an application submitted to the Secretary. Not later than 75 days after the date of the receipt of the application, the Secretary shall issue an order approving or denying the application.''

Subsec. (m)(4). Pub. L. 105–115, § 203(2)(B), inserted at end ''In a case described in subparagraph (B) in which a physician uses a device without an approval from an institutional review committee, the physician shall, after the use of the device, notify the chairperson of the local institutional review committee of such use. Such notification shall include the identification of the patient involved, the date on which the device was used, and the reason for the use.''

Subsec. (m)(4)(B). Pub. L. 105–115, § 203(2)(A), inserted before period at end '', unless a physician determines in an emergency situation that approval from a local institutional review committee can not be obtained in time to prevent serious harm or death to a patient''.

Subsec. (m)(5). Pub. L. 105–115, § 203(3), amended par. (5) generally. Prior to amendment, par. (5) read as follows: ''An exemption under paragraph (2) shall be for a term of 18 months and may only be initially granted in the 5-year period beginning on the date regulations under paragraph (6) take effect. The Secretary may extend such an exemption for a period of 18 months if the Secretary is able to make the findings set forth in paragraph (2) and if the applicant supplies information demonstrating compliance with paragraph (3). An exemption may be extended more than once and may be extended after the expiration of such 5-year period.''

Subsec. (m)(6). Pub. L. 105–115, § 203(4), amended par. (6) generally. Prior to amendment, par. (6) read as follows: ''Within one year of November 28, 1990, the Secretary shall issue regulations to implement this subsection.''

1992—Subsec. (g)(2)(A). Pub. L. 102–571 substituted ''379e'' for ''376''.

1990—Subsec. (c). Pub. L. 101–629, § 11(1), substituted ''from class III to class II or class I'' for ''under section 360c of this title from class III to class II'' and inserted ''(1) in accordance with subsection (h) of this section, and (2)'' after ''except''.

Subsec. (f)(1)(A). Pub. L. 101–629, § 18(e), inserted ''preproduction design validation (including a process to assess the performance of a device but not including an evaluation of the safety or effectiveness of a device),'' after ''manufacture,''.

Subsec. (h)(3). Pub. L. 101–629, § 11(2)(A), substituted ''Except as provided in paragraph (4), any'' for ''Any''.

Subsec. (h)(4). Pub. L. 101–629, § 11(2)(B), added par. (4).

Subsec. (i). Pub. L. 101–629, § 6(b)(2), substituted ''section 360d(b)(5)(B)'' for ''section 360d(g)(5)(B)''.

Subsec. (j). Pub. L. 101–629, § 3(b)(2), substituted ''Except as provided in section 360i(e) of this title, no'' for ''No''.

Subsec. (*l*)(2). Pub. L. 101–629, § 18(f), struck out ''and after affording the petitioner an opportunity for an informal hearing'' after ''under this paragraph''.

Pub. L. 101–629, § 5(c)(2), substituted ''The Secretary may initiate the reclassification of a device classified into class III under paragraph (1) of this subsection or the manufacturer'' for ''The manufacturer''.

Subsec. (*l*)(5). Pub. L. 101–629, § 4(b)(2), added par. (5).

Subsec. (m). Pub. L. 101–629, § 14(a), added subsec. (m).

EFFECTIVE DATE OF 1997 AMENDMENT

Amendment by sections 201(a), 203, 216(a)(1), and 410(a) of Pub. L. 105–115 effective 90 days after Nov. 21, 1997, except as otherwise provided, see section 501 of Pub. L. 105–115, set out as a note under section 321 of this title.

EFFECTIVE DATE OF 1990 AMENDMENT

Section 14(b) of Pub. L. 101–629 provided that: ''Subsection (m) of section 520 of the Federal Food, Drug, and Cosmetic Act [21 U.S.C. 360j(m)], as added by the amendment made by subsection (a), shall take effect on the effective date of the regulations issued by the Secretary under paragraph (6) of such subsection.''

GUIDANCE

Pub. L. 110–85, title III, § 303(c), Sept. 27, 2007, 121 Stat. 862, provided that: ''Not later than 180 days after the date of the enactment of this Act [Sept. 27, 2007], the Commissioner of Food and Drugs shall issue guidance for institutional review committees on how to evaluate requests for approval for devices for which a humanitarian device exemption under section 520(m)(2) of the Federal Food, Drug, and Cosmetic Act (21 U.S.C. 360j(m)(2)) has been granted.''

Pub. L. 107–250, title II, § 213, Oct. 26, 2002, 116 Stat. 1614, provided that: ''Not later than 270 days after the date of the enactment of this Act [Oct. 26, 2002], the Secretary of Health and Human Services shall issue guidance on the following:

''(1) The type of information necessary to provide reasonable assurance of the safety and effectiveness of medical devices intended for use in pediatric populations.

''(2) Protections for pediatric subjects in clinical investigations of the safety or effectiveness of such devices.''

REPORT ON HUMANITARIAN DEVICE EXEMPTIONS

Section 14(c) of Pub. L. 101–629 directed Secretary of Health and Human Services, within 4 years after issuance of regulations under 21 U.S.C. 360j(m)(6), to report to Congress on types of devices exempted, an evaluation of effects of such section, and a recommendation on extension of the section.

REFERENCES IN OTHER LAWS TO GS–16, 17, OR 18 PAY RATES

References in laws to the rates of pay for GS–16, 17, or 18, or to maximum rates of pay under the General Schedule, to be considered references to rates payable under specified sections of Title 5, Government Organization and Employees, see section 529 [title I, § 101(c)(1)] of Pub. L. 101–509, set out in a note under section 5376 of Title 5.

§ 360k. State and local requirements respecting devices

(a) General rule

Except as provided in subsection (b) of this section, no State or political subdivision of a State may establish or continue in effect with

respect to a device intended for human use any requirement—

(1) which is different from, or in addition to, any requirement applicable under this chapter to the device, and

(2) which relates to the safety or effectiveness of the device or to any other matter included in a requirement applicable to the device under this chapter.

(b) Exempt requirements

Upon application of a State or a political subdivision thereof, the Secretary may, by regulation promulgated after notice and opportunity for an oral hearing, exempt from subsection (a) of this section, under such conditions as may be prescribed in such regulation, a requirement of such State or political subdivision applicable to a device intended for human use if—

(1) the requirement is more stringent than a requirement under this chapter which would be applicable to the device if an exemption were not in effect under this subsection; or

(2) the requirement—

(A) is required by compelling local conditions, and

(B) compliance with the requirement would not cause the device to be in violation of any applicable requirement under this chapter.

(June 25, 1938, ch. 675, § 521, as added Pub. L. 94–295, § 2, May 28, 1976, 90 Stat. 574.)

§ 360*l*. Postmarket surveillance

(a) Postmarket surveillance

(1) In general

(A) Conduct

The Secretary may by order require a manufacturer to conduct postmarket surveillance for any device of the manufacturer that is a class II or class III device—

(i) the failure of which would be reasonably likely to have serious adverse health consequences;

(ii) that is expected to have significant use in pediatric populations; or

(iii) that is intended to be—

(I) implanted in the human body for more than 1 year; or

(II) a life-sustaining or life-supporting device used outside a device user facility.

(B) Condition

The Secretary may order a postmarket surveillance under subparagraph (A) as a condition to approval or clearance of a device described in subparagraph (A)(ii).

(2) Rule of construction

The provisions of paragraph (1) shall have no effect on authorities otherwise provided under the[1] chapter or regulations issued under this chapter.

(b) Surveillance approval

(1) In general

Each manufacturer required to conduct a surveillance of a device shall, within 30 days of receiving an order from the Secretary prescribing that the manufacturer is required under this section to conduct such surveillance, submit, for the approval of the Secretary, a plan for the required surveillance. The Secretary, within 60 days of the receipt of such plan, shall determine if the person designated to conduct the surveillance has appropriate qualifications and experience to undertake such surveillance and if the plan will result in the collection of useful data that can reveal unforeseen adverse events or other information necessary to protect the public health. Except as provided in paragraph (2), the Secretary, in consultation with the manufacturer, may by order require a prospective surveillance period of up to 36 months. Except as provided in paragraph (2), any determination by the Secretary that a longer period is necessary shall be made by mutual agreement between the Secretary and the manufacturer or, if no agreement can be reached, after the completion of a dispute resolution process as described in section 360bbb–1 of this title.

(2) Longer surveillance for pediatric devices

The Secretary may by order require a prospective surveillance period of more than 36 months with respect to a device that is expected to have significant use in pediatric populations if such period of more than 36 months is necessary in order to assess the impact of the device on growth and development, or the effects of growth, development, activity level, or other factors on the safety or efficacy of the device.

(c) Dispute resolution

A manufacturer may request review under section 360bbb–1 of this title of any order or condition requiring postmarket surveillance under this section. During the pendency of such review, the device subject to such a postmarket surveillance order or condition shall not, because of noncompliance with such order or condition, be deemed in violation of section 331(q)(1)(C) of this title, adulterated under section 351(f)(1) of this title, misbranded under section 352(t)(3) of this title, or in violation of, as applicable, section 360(k) of this title or section 360e of this title, unless deemed necessary to protect the public health.

(June 25, 1938, ch. 675, § 522, as added Pub. L. 101–629, § 10, Nov. 28, 1990, 104 Stat. 4521; amended Pub. L. 102–300, § 3(b), June 16, 1992, 106 Stat. 239; Pub. L. 105–115, title II, § 212, Nov. 21, 1997, 111 Stat. 2346; Pub. L. 110–85, title III, § 307, Sept. 27, 2007, 121 Stat. 865.)

AMENDMENTS

2007—Pub. L. 110–85, § 307(1), made technical amendment to section catchline.

Subsec. (a). Pub. L. 110–85, § 307(2), added subsec. (a) and struck out former subsec. (a). Prior to amendment, text read as follows: ''The Secretary may by order require a manufacturer to conduct postmarket surveillance for any device of the manufacturer which is a class II or class III device the failure of which would be reasonably likely to have serious adverse health consequences or which is intended to be—

''(1) implanted in the human body for more than one year, or

''(2) a life sustaining or life supporting device used outside a device user facility.''

[1] So in original. Probably should be ''this''.