

**§ 493.807**

as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA.

(b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part.

(c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists:

(1) There is immediate jeopardy to patient health and safety.

(2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance.

(3) The laboratory has a poor compliance history.

[57 FR 7146, Feb. 28, 1992, as amended at 60 FR 20048, Apr. 24, 1995; 63 FR 26737, May 14, 1998; 68 FR 3702, Jan. 24, 2003]

**§ 493.807 Condition: Reinstatement of laboratories performing nonwaived testing.**

(a) If a laboratory's certificate is suspended or limited or its Medicare or Medicaid approval is cancelled or its Medicare or Medicaid payments are suspended because it fails to participate successfully in proficiency testing for one or more specialties, subspecialties, analyte or test, or voluntarily withdraws its certification under CLIA for the failed specialty, subspecialty, or analyte, the laboratory must then demonstrate sustained satisfactory performance on two consecutive proficiency testing events, one of which may be on site, before CMS will consider it for reinstatement for certification and Medicare or Medicaid ap-

**42 CFR Ch. IV (10-1-10 Edition)**

proval in that specialty, subspecialty, analyte or test.

(b) The cancellation period for Medicare and Medicaid approval or period for suspension of Medicare or Medicaid payments or suspension or limitation of certification under CLIA for the failed specialty, subspecialty, or analyte or test is for a period of not less than six months from the date of cancellation, limitation or suspension of the CLIA certificate.

[58 FR 5228, Jan. 19, 1993, as amended at 60 FR 20048, Apr. 24, 1995]

**PROFICIENCY TESTING BY SPECIALTY AND SUBSPECIALTY FOR LABORATORIES PERFORMING TESTS OF MODERATE COMPLEXITY (INCLUDING THE SUBCATEGORY), HIGH COMPLEXITY, OR ANY COMBINATION OF THESE TESTS**

**§ 493.821 Condition: Microbiology.**

The specialty of microbiology includes, for purposes of proficiency testing, the subspecialties of bacteriology, mycobacteriology, mycology, parasitology and virology.

**§ 493.823 Standard: Bacteriology.**

(a) Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.

(b) Failure to participate in a testing event is unsatisfactory performance and results in a score of 0 for the testing event. Consideration may be given to those laboratories failing to participate in a testing event only if—

(1) Patient testing was suspended during the time frame allotted for testing and reporting proficiency testing results;

(2) The laboratory notifies the inspecting agency and the proficiency testing program within the time frame for submitting proficiency testing results of the suspension of patient testing and the circumstances associated with failure to perform tests on proficiency testing samples; and

(3) The laboratory participated in the previous two proficiency testing events.

(c) Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory

performance and results in a score of 0 for the testing event.

(d)(1) For any unsatisfactory testing event for reasons other than a failure to participate, the laboratory must undertake appropriate training and employ the technical assistance necessary to correct problems associated with a proficiency testing failure.

(2) Remedial action must be taken and documented, and the documentation must be maintained by the laboratory for two years from the date of participation in the proficiency testing event.

(e) Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

**§ 493.825 Standard; Mycobacteriology.**

(a) Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.

(b) Failure to participate in a testing event is unsatisfactory performance and results in a score of 0 for the testing event. Consideration may be given to those laboratories failing to participate in a testing event only if—

(1) Patient testing was suspended during the time frame allotted for testing and reporting proficiency testing results;

(2) The laboratory notifies the inspecting agency and the proficiency testing program within the time frame for submitting proficiency testing results of the suspension of patient testing and the circumstances associated with failure to perform tests on proficiency testing samples; and

(3) The laboratory participated in the previous two proficiency testing events.

(c) Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory performance and results in a score of 0 for the testing event.

(d)(1) For any unsatisfactory testing event for reasons other than a failure to participate, the laboratory must undertake appropriate training and employ the technical assistance necessary to correct problems associated with a proficiency testing failure.

(2) Remedial action must be taken and documented, and the documentation must be maintained by the laboratory for two years from the date of participation in the proficiency testing event.

(e) Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

**§ 493.827 Standard; Mycology.**

(a) Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.

(b) Failure to participate in a testing event is unsatisfactory performance and results in a score of 0 for the testing event. Consideration may be given to those laboratories failing to participate in a testing event only if—

(1) Patient testing was suspended during the time frame allotted for testing and reporting proficiency testing results;

(2) The laboratory notifies the inspecting agency and the proficiency testing program within the time frame for submitting proficiency testing results of the suspension of patient testing and the circumstances associated with failure to perform tests on proficiency testing samples; and

(3) The laboratory participated in the previous two proficiency testing events.

(c) Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory performance and results in a score of 0 for the testing event.

(d)(1) For any unsatisfactory testing event for reasons other than a failure to participate, the laboratory must undertake appropriate training and employ the technical assistance necessary to correct problems associated with a proficiency testing failure.

(2) Remedial action must be taken and documented, and the documentation must be maintained by the laboratory for two years from the date of participation in the proficiency testing event.

(e) Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events