

§ 660.36

PRODUCTS DERIVED FROM HUMAN BLOOD WILL NOT TRANSMIT INFECTIOUS AGENTS.”

(l) The package insert or the antigenic constitution matrix for each lot of product shall specify the date of manufacture or the length of the dating period.

(m) Manufacturers shall identify with a permanent donor code in the product labeling each donor of peripheral blood used for detection or identification of unexpected antibodies.

[52 FR 37450, Oct. 7, 1987, as amended at 67 FR 9587, Mar. 4, 2002]

§ 660.36 Samples and protocols.

(a) The following shall be submitted to the Center for Biologics Evaluation and Research Sample Custodian (ATTN: HFM-672) (see mailing addresses in § 600.2 of this chapter), within 30 days after each routine establishment inspection by FDA.

(1) From a lot of final product, samples from a cell panel intended for identification of unexpected antibodies. The sample shall be packaged as for distribution and shall have at least 14 days remaining in the dating period when shipped to the Center for Biologics Evaluation and Research.

(2) A protocol which shall include the following:

(i) Complete test records of at least two donors of the samples submitted, including original and confirmation phenotyping records.

(ii) Bleeding records or receipt records which indicate collection date, volume, and HBsAg test results.

(iii) Manufacturing records which document all steps involved in the preparation of the product.

(iv) Test results which verify that the final product meets specifications.

(v) Identity test results.

(b) A copy of the antigenic constitution matrix specifying the antigens present or absent shall be submitted to the Director, Center for Biologics Evaluation and Research, at the time of initial distribution of each lot of Reagent Red Blood Cells for detection or identification of unexpected antibodies. Products designed exclusively to identify Anti-A, Anti-A₁, and Anti-B, as well as products composed entirely of

21 CFR Ch. I (4–1–10 Edition)

umbilical cord cells, are excluded from this requirement.

(c) Except for umbilical cord samples, whenever a new donor is used, a sample of red blood cells from each new donor used in a cell panel intended for the identification of unexpected antibodies shall be submitted by the manufacturer to the Director, Center for Biologics Evaluation and Research. The sample should contain a minimum volume of 0.5 milliliter of red blood cells.

[52 FR 37450, Oct. 7, 1987, as amended at 55 FR 11013 and 11015, Mar. 26, 1990; 67 FR 9587, Mar. 4, 2002; 70 FR 14985, Mar. 24, 2005]

Subpart E—Hepatitis B Surface Antigen

SOURCE: 44 FR 36382, June 22, 1979, unless otherwise noted.

§ 660.40 Hepatitis B Surface Antigen.

(a) *Proper name and definition.* The proper name of this product shall be Hepatitis B Surface Antigen (HBsAg), which shall consist of a serum or tissue preparation containing one or more subtypes of the Hepatitis B Surface Antigen.

(b) *Source.* The source of the product shall be blood, plasma, serum, or tissue, obtained aseptically from nonhuman primates that have met the applicable requirements of § 600.11 of this chapter, or from human donors whose blood is positive for the Hepatitis B Surface Antigen.

§ 660.41 Processing.

(a) *Method.* The processing method shall be one that has been shown to yield consistently a specific and potent final product, free of properties which would adversely affect the test results when the product is tested by the methods recommended by the manufacturer in the package insert. The product and all ancillary reagents and materials supplied in the package with the product shall be manufactured in a manner that will reduce the risk of transmitting type B viral hepatitis.

(b) *Ancillary reagents and materials.* All ancillary reagents and materials supplied in the package with the product shall meet generally accepted

standards of purity and quality and shall be effectively segregated and otherwise manufactured in a manner that will reduce the risk of contaminating the product and other biological products. Ancillary reagents and materials accompanying the product, which are used in the performance of the test as described by the manufacturer's recommended test procedures, shall have been shown not to affect adversely the product within the prescribed dating period.

(c) *Final container.* A final container shall be sufficiently transparent to permit visual inspection of the contents for presence of particulate matter and increased turbidity. The effectiveness of the contents of a final container shall be maintained throughout its dating period.

(d) *Date of manufacture.* The date of manufacture of Hepatitis B Surface Antigen that has been iodinated with radioactive iodine (¹²⁵I) shall be the day of labeling the antibody with the radionuclide.

[44 FR 36382, June 22, 1979, as amended at 49 FR 1685, Jan. 13, 1984]

§ 660.43 Potency test.

To be satisfactory for release, each filling of Hepatitis B Surface Antigen shall be tested against the Reference Hepatitis B Antiserum Panel and shall be sufficiently potent to be able to detect the antibody in the appropriate sera of the reference panel by all test methods recommended by the manufacturer in the package insert.

§ 660.44 Specificity.

Each filling of the product shall be specific for Hepatitis B Surface Antigen as determined by specificity tests found acceptable to the Director, Center for Biologics Evaluation and Research.

[44 FR 36382, June 22, 1979, as amended at 49 FR 23834, June 8, 1984; 55 FR 11013, Mar. 26, 1990]

§ 660.45 Labeling.

In addition to the requirements of §§ 610.60, 610.61, and 809.10 of this chapter, the labeling shall bear the following:

(a) The "d and y" antigen subtype and the source of the product to follow immediately the proper name on both the final container label and the package label. If the product is intended to identify antibodies to the "r and w" antigen subtype, the antigen subtype designation shall include the "r and w" antigen subtype.

(b) The name of the test method(s) recommended for use of the product on the package label and on the final container label, when capable of bearing a full label (see § 610.60(a) of this chapter).

(c) A warning on the package label and on the final container label stating that the product is capable of transmitting hepatitis and should be handled accordingly.

(d) The package shall include a package insert providing (1) detailed instructions for use, (2) an adequate description of all recommended test methods, and (3) warnings as to possible hazards, including hepatitis transmitted in handling the product and any ancillary reagents and materials accompanying the product.

§ 660.46 Samples; protocols; official release.

(a) *Samples.* (1) For the purposes of this section, a sample of product not iodinated with ¹²⁵I means a sample from each filling of each lot packaged as for distribution, including all ancillary reagents and materials; and a sample of product iodinated with ¹²⁵I or unlyophilized HBsAg-coated red blood cells means a sample from each lot of diagnostic test kits in a finished package, including all ancillary reagents and materials.

(2) Unless the Director, Center for Biologics Evaluation and Research, determines that the reliability and consistency of the finished product can be assured with a smaller quantity of sample or no sample and specifically reduces or eliminates the required quantity of sample, each manufacturer shall submit the following samples to the Director, Center for Biologics Evaluation and Research (see mailing addresses in § 600.2 of this chapter), within 5 working days after the manufacturer has satisfactorily completed all tests on the samples: