§640.6

(d) *Sterility test*. Whole Blood intended for transfusion shall not be tested for sterility by a method that entails entering the final container before the blood is used for transfusion.

(e) *Inspection*. Whole Blood shall be inspected visually during storage and immediately prior to issue. If the color or physical appearance is abnormal or there is any indication or suspicion of microbial contamination the unit of Whole Blood shall not be issued for transfusion.

(f) Test for communicable disease agents. Whole Blood shall be tested for evidence of infection due to communicable disease agents as required under § 610.40 of this chapter.

[38 FR 32089, Nov. 20, 1973, as amended at 50 FR 4138, Jan. 29, 1985; 53 FR 117, Jan. 5, 1988; 53 FR 12764, Apr. 19, 1988; 64 FR 45372, Aug. 19, 1999; 66 FR 1836, Jan. 10, 2001; 66 FR 31165, June 11, 2001; 66 FR 40889, Aug. 6, 2001]

§640.6 Modifications of Whole Blood.

Upon approval by the Director, Center for Biologics Evaluation and Research, of a supplement to the biologics license application for Whole Blood a manufacturer may prepare Whole Blood from which the antihemophilic factor has been removed, provided the Whole Blood meets the applicable requirements of this subchapter and the following conditions are met:

(a) The antihemophilic factor shall be removed in accordance with paragraphs (a), (b), and (c) of §640.52.

(b) Although the closed system between the red blood cells and plasma shall be maintained, the red blood cells shall be maintained between 1 and 6 °C at all times, including that time when the plasma is being frozen for removal of the antihemophilic factor.

[38 FR 32089, Nov. 20, 1973, as amended at 49
FR 23834, June 8, 1984; 50 FR 4138, Jan. 29, 1985; 55 FR 11013, Mar. 26, 1990; 59 FR 49351, Sept. 28, 1994; 64 FR 45372, Aug. 19, 1999; 64 FR 56453, Oct. 20, 1999]

Subpart B—Red Blood Cells

§640.10 Red Blood Cells.

The proper name of this product shall be Red Blood Cells. The product is de-

21 CFR Ch. I (4–1–12 Edition)

fined as red blood cells remaining after separating plasma from human blood.

[38 FR 32089, Nov. 20, 1973, as amended at 50 FR 4138, Jan. 29, 1985]

§640.11 General requirements.

(a) Storage. Immediately after processing, the Red Blood Cells shall be placed in storage and maintained at a temperature between 1 and 6 $^{\circ}$ C.

(b) Inspection. The product shall be inspected immediately after separation of the plasma, periodically during storage, and at the time of issue. The product shall not be issued if there is any abnormality in color or physical appearance or if there is any indication of microbial contamination.

[38 FR 32089, Nov. 20, 1973, as amended at 41 FR 18292, May 3, 1976; 42 FR 59878, Nov. 11, 1977; 50 FR 4139, Jan. 29, 1985]

§640.12 Suitability of donor.

The source blood for Red Blood Cells shall be obtained from a donor who meets the criteria for donor suitability prescribed in § 640.3.

[38 FR 32089, Nov. 20, 1973, as amended at 50 FR 4139, Jan. 29, 1985]

§640.13 Collection of the blood.

(a) The source blood shall be collected as prescribed in §640.4.

(b) Source blood may also be derived from Whole Blood manufactured in accordance with applicable provisions of this subchapter.

[38 FR 32089, Nov. 20, 1973, as amended at 50 FR 4139, Jan. 29, 1985; 64 FR 45372, Aug. 19, 1999]

§640.14 Testing the blood.

Blood from which Red Blood Cells are prepared shall be tested as prescribed in §610.40 of this chapter and §640.5 (a), (b), and (c).

 $[53\ {\rm FR}\ 117,\ {\rm Jan.}\ 5,\ 1988,\ {\rm as}\ {\rm amended}\ {\rm at}\ 66\ {\rm FR}\ 31165,\ {\rm June}\ 11,\ 2001]$

§640.15 Segments for testing.

Segments collected in integral tubing shall meet the following standards:

(a) One or more segments shall be provided with each unit of Whole Blood or Red Blood Cells when issued or reissued.