

§ 493.1253

42 CFR Ch. IV (10–1–14 Edition)

(4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports.

(c) Reagents, solutions, culture media, control materials, calibration materials, and other supplies, as appropriate, must be labeled to indicate the following:

(1) Identity and when significant, titer, strength or concentration.

(2) Storage requirements.

(3) Preparation and expiration dates.

(4) Other pertinent information required for proper use.

(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.

(e) Components of reagent kits of different lot numbers must not be interchanged unless otherwise specified by the manufacturer.

**§ 493.1253 Standard: Establishment and verification of performance specifications.**

(a) *Applicability.* Laboratories are not required to verify or establish performance specifications for any test system used by the laboratory before April 24, 2003.

(b)(1) *Verification of performance specifications.* Each laboratory that introduces an unmodified, FDA-cleared or approved test system must do the following before reporting patient test results:

(i) Demonstrate that it can obtain performance specifications comparable to those established by the manufacturer for the following performance characteristics:

(A) Accuracy.

(B) Precision.

(C) Reportable range of test results for the test system.

(ii) Verify that the manufacturer's reference intervals (normal values) are appropriate for the laboratory's patient population.

(2) *Establishment of performance specifications.* Each laboratory that modifies an FDA-cleared or approved test system, or introduces a test system not subject to FDA clearance or approval

(including methods developed in-house and standardized methods such as text book procedures), or uses a test system in which performance specifications are not provided by the manufacturer must, before reporting patient test results, establish for each test system the performance specifications for the following performance characteristics, as applicable:

(i) Accuracy.

(ii) Precision.

(iii) Analytical sensitivity.

(iv) Analytical specificity to include interfering substances.

(v) Reportable range of test results for the test system.

(vi) Reference intervals (normal values).

(vii) Any other performance characteristic required for test performance.

(3) *Determination of calibration and control procedures.* The laboratory must determine the test system's calibration procedures and control procedures based upon the performance specifications verified or established under paragraph (b)(1) or (b)(2) of this section.

(c) *Documentation.* The laboratory must document all activities specified in this section.

[68 FR 3703, Jan. 24, 2003; 68 FR 50724, Aug. 22, 2003]

**§ 493.1254 Standard: Maintenance and function checks.**

(a) *Unmodified manufacturer's equipment, instruments, or test systems.* The laboratory must perform and document the following:

(1) Maintenance as defined by the manufacturer and with at least the frequency specified by the manufacturer.

(2) Function checks as defined by the manufacturer and with at least the frequency specified by the manufacturer. Function checks must be within the manufacturer's established limits before patient testing is conducted.

(b) *Equipment, instruments, or test systems developed in-house, commercially available and modified by the laboratory, or maintenance and function check protocols are not provided by the manufacturer.* The laboratory must do the following: